

Cobb Sizing Testers



Manual Clamp



Hand Roll



Automatic Clamp



INSTRUMENT SIZE

Manual Clamping

Depth: 140 mm (5.5 in.)

Height: 80 mm (3.0 in.)

Width: 140 mm (5.5 in.)

Automatic Clamping

Depth: 200 mm (8.0 in.)

Height: 180 mm (7.0 in.)

Width: 200 mm (8.0 in.)

WEIGHT

2 kg (4.4 lb)

CONNECTIONS

None

ORDERING INFORMATION

Automatic Clamping

Used for	Catalog number
Paper	61-05-00-0002
Board	60-05-00-0001

Manual Clamping

Size(s)	Catalog number
100 sq. cm.	61-04-00-0002
10, 25, 100 sq. cm	61-04-00-0001

Hand Roll	61-04-01
------------------	----------

INTRODUCTION

Water absorptiveness (Cobb value) is the mass of water absorbed in a specific time by a 1 sq. meter sample of paper, board or corrugated under conditions specified in standards such as TAPPI T 441. Cobb sizing testers provide a simple, standardized method for fast determination of the quantity of water absorbed by a paper or board in a given time.

APPLICATIONS

Paper, board, corrugated

REQUIRED ACCESSORIES

Stainless Steel Hand Roll (for compliance with recognized standards)

FEATURES

- Single-handed automatic clamping and release
- Manual clamping and sample positioning

The TMI Group of Companies



www.testingmachines.com

Testing Machines Inc. 2 Fleetwood Court, Ronkonkoma, NY 11779 USA
Tel: 800-678-3221 (USA) or (631) 439-5400, Fax: (631) 439-5420, e-mail: info@testingmachines.com

Messmer Instruments Ltd Unit F1 Imperial Business Estate, West Mill, Gravesend, Kent DA11 0DL UK
Tel: +44 (0)1474 566488, Fax: +44 (0)1474 560310, e-mail: messmer@testingmachines.com

Büchel BV Fokkerstraat 24, 3905 KV Veenendaal, Netherlands
Tel: +31 (0)318 521500, Fax: +31 (0)318 540358, e-mail: buchel@testingmachines.com

Adamel Lhomargy SAS Z.A. de l'Habitat, Batiment 6, Route d'Ozoir, 77680, Roissy en Brie, France
Tel: +33 (0)1 64409210, Fax: +33 (0)1 64409211, e-mail: adamellhomargy@testingmachines.com

TMI Canada 6 Ronald Drive, Montreal West, Quebec H4X 1M8
Tel: (514) 489-8491, Fax (514) 489-6044, e-mail: canada@testingmachines.com